

CCT ROOMS

- Ensure each bed in Rooms 1-5 has a working Bag Valve Mask
- Ensure each bed in Rooms 1-5 has working suction (canister to tube to yankauer).
- Ensure Ultrasounds are plugged in near trauma bay
- Ensure Glidescope is loaded and functioning near trauma bay.
- Ensure you have the equipment needed to quickly draw a full set of labs and put in an IV.
- The bougies are now in the Pyxis, you will need to order them in order to get one
- Lastly, know how to set up the Level 1 infusers. These are not being prepared before hand but will be set up when indicated.

TRAUMA BAY 1 – RED AIRWAY CARTS

- Ensure top shelf contains size 6 to 9 ET tubes with 2 stylets. Please don't put stylets into the ET tubes before use.
- Make sure each cart has 2 handles, 2 Mac blades, 2 Miller blades. Please test to make sure these work (especially the light).
- Ensure both airway carts have working BVMs & suction
- Ensure the drawers below the top drawer are reasonable (nursing will stock, check though).

BLUE / WHITE AIRWAY CARTS

- They are mobile with one typically placed near Rooms 2-3 and another placed near Room 5.
- These carts are to be organized, stocked, and maintained **by the residents.**

- There are three (3) drawers on these carts arranged in the following manner:
- **Drawer 1:**
 - Scalpels (11 blades)
 - Tongue blades
 - ETCO2 detectors
 - Silk tape
 - Laryngoscope handles
 - Miller 3 and 4 blades
 - Mac 3 and 4 blades
 - ET tubes from sizes 6-9
- **Drawer 2:**
 - Nasal cannulas
 - Non-rebreather masks
 - LMA (size 3.5 and 4.5)
 - Nasopharyngeal airways
 - Lubricating jelly
 - Oropharyngeal airways
 - Betadine bottle
 - Benzoin bottle
- **Drawer 3:**
 - ETT fastener devices
 - Extra stylets
 - 20cc syringes
 - Extra unopened endotracheal tubes.

If you are missing any supplies in the airway cart go to the CCT Charge Nurse

We would like to thank one of our alumni and previous chiefs, Ben Kaufman, for creating this document.